

# **MAPS**

## **Malt Analytes Proficiency Testing Scheme**

## **Scheme Description**

## LGC Proficiency Testing

1 Chamberhall Business Park Chamberhall Green Bury Lancashire BL9 0AP United Kingdom

Telephone: +44 (0) 161 762 2500 Email: axiopt@lgcgroup.com Website www.lgcstandards.com





#### Record of issue status and modifications

ISSUE	ISSUE DATE	DETAILS	AUTHORISED BY
11	Sept 2013	Details for sample 3 updated. Methods for FAN and EBC reject fraction updated for 2014.	W.Gaunt
12	Jan 2014	Assigned value text updated. Methods updated for various parameters for 2014.	W.Gaunt
13	Aug 2014	Methods updated for various parameters.	W.Gaunt
14	Sep 2014	SDPA updated for colour in sample 4. Inclusion of subcontracting information in 'Test Materials' section.	W.Gaunt
15	Sep 2015	Black and crystal malt separated (now 4A and 4B) Removed Hard copy report information.	W.Gaunt A.McCarthy
16	Sept 2016	Details updated for TSN (EBC & IOB) and EBC fermentability	W.Gaunt
17	Sept 2017	IOB fermentability (unboiled 0.2) and IOB fermentability (boiled) removed for 2018. Kjeldahl and Dumas methods for Kolbach index deleted and replaced with 'Calculation'. Total UV digestion added to TSN (EBC Wort) & TSN (IoB wort). SDPA for EBC fermentability reduced to 1.0. New samples added for high DP (6) and malted wheat (7).	W.Gaunt
18	Oct 2017	Analytes amended for samples 6 and 7. Methods updated for Beta glucan (EBC & IOB)	W.Gaunt
19	Sep 2018	Analyte clarification for 0.2mm and FAN in malted wheat	W.Gaunt
20	Aug 2019	DPWK and DP (IoB) renamed, SDPA updated for DMSP (malt) SDPA revised for DMSP (malt)	W.Gaunt
21	July 2021	Updated email address and UKAS logo	A Collins
23	Sept 2022	Water sensitivity included for barley. Saccharification time added for malt.	W.Gaunt
24	Jan 2023	SDPA updated for colour in black and crystal malts	W.Gaunt

#### Notes:

Where this document has been translated, the English version shall remain the definitive version

#### **Scheme Aims and Organisation**

The primary aim of the Malt Analytes Proficiency Testing Scheme (MAPS) is to enable laboratories performing the analysis of malt and barley to monitor their performance and compare it with that of their peers. MAPS also aims to provide information to participants on technical issues and methodologies relating to testing of malt and barley.

The MAPS scheme year operates from January to December. Further information about MAPS, including test material availability, round despatch dates and reporting deadlines, are available on the current MAPS application form.

#### **Test Materials**

Details of test materials available in MAPS are given in Appendix A. The test parameters are continually reviewed to ensure they meet the needs of current laboratory testing and regulatory requirements.

Test material batches are tested for homogeneity for at least one test parameter where deemed appropriate. Details of homogeneity tests performed and results are given in the MAPS Scheme Reports.

Some aspects of the scheme, such as test material production, homogeneity testing and stability assessment, can from time to time be subcontracted. When subcontracting occurs, it is placed with a competent subcontractor and LGC is responsible for this work. The planning of the scheme, the evaluation of performance and the authorisation of the final report will never be subcontracted.

#### **Statistical Analysis**

Information on the statistics used in MAPS can be found in the General Protocol and in the Scheme Report. Methods for determining assigned values and the values for SDPA used for individual samples are given in Appendix A

#### **Methods**

Methods are listed in PORTAL. Please select the most appropriate method from the list. If none of the methods are appropriate, then please report your method as 'Other' and record a brief description in the Comments Section in PORTAL.

#### **Results and Reports**

MAPS results are returned through our electronic reporting software, PORTAL, full instructions for which are provided by email.

MAPS reports will be available on the website within 10 working days of round closure. Participants will be emailed a link to the report when it is available.

#### **APPENDIX A - Description of abbreviations used**

Assigned Value (AV) - The assigned value may be derived in the following ways:

From the robust mean (RMean). This is the median of participant results after the removal of test results that are inappropriate for statistical evaluation, e.g. miscalculations, transpositions and other gross errors. Generally, the assigned value will be set using results from all methods, unless the measurement is considered method-dependant, in which case the assigned value will be set by method as illustrated in the report tables.

For some analytes, where there is a recognised reference method for that type of measurement, this may be used as the assigned value for a particular analyte i.e. it would be applied to results obtained by any method.

Traceability: Assigned values which are derived from the participant results, or a sub-set of the results are not traceable to an international measurement standard. The uncertainty of assigned values derived in this way is estimated from the participant results, according to ISO 13528.

 From a formulation value (Formulation). This denotes the use of an assigned value derived from sample preparation details, where known and exact quantities of analyte have been used to prepare the sample.

Traceability: Assigned values calculated from the formulation of the test sample are traceable, via an unbroken metrological traceability chain, to an international measurement standard. The measurement uncertainty of the assigned value is calculated using the contributions from each calibration in the traceability chain.

• From a qualitative formulation (Qual Form). This applies to qualitative tests where the assigned value is simply based on the presence/absence of the analyte in the test material.

Traceability: Assigned values calculated from the qualitative formulation of the test sample are traceable to a certified reference standard or a microbiological reference strain.

 From expert labs (Expert). The assigned value for the analyte is provided by an 'expert' laboratory.

Traceability: Assigned values provided by an 'expert' laboratory may be traceable to an international measurement standard, according to the laboratory and the method used. The uncertainty of measurement for an assigned value produced in this way will be provided by the laboratory undertaking the analysis. Details of traceability and the associated uncertainty will be provided in the report for the scheme/round.

#### Range

This indicates the concentration range at which the analyte may be present in the test material.

#### SDPA

SDPA represents the 'standard deviation for proficiency assessment' which is used to assess participant performance for the measurement of each analyte. This may be a fixed value (as stated), a percentage (%) of the assigned value or based on the robust standard deviation of the participant measurement results, either across all methods or by method depending on whether the measurement made is method dependent (see assigned value).

#### **Units**

This indicates the units used for the assessment of data. These are the units in which participants should report their results. For some analytes in some schemes participants may have a choice of which units to report their results, however, the units stipulated in this scheme description are the default units to which any results reported using allowable alternative results will be converted to.

#### DP

This indicates the number of decimal places to which participants should report their measurement results.

Sample PT-MP-01: Brewers and Distiller malt for chemical analysis

Participants will receive 2 x 1000ml pots of malt

## 1) Malt sample - Malt dependent tests

Analyte	Method	Range	AV	SDPA	Units	DP
Moisture	Oven Method NIR (Near Infrared)	3 to 10	RMean	0.15	% w/w	1
Diastatic Power (EBC)	Iodometric (reference method) Segmented flow (e.g. Skalar) Calculated value (from IOB)	≤350 350-500 >500	RMean (Iodometric)	20 25 30	WK units, dry basis	0
Diastatic Power (IoB)	Fehling's Ferricyanide Segmented flow (e.g. Skalar) Calculated value (from DPWK)	- ≤91 - 91-128 - >128	Calculated value	9.1 10.4 11.6	IoB, as is basis	0
Alpha Amylase	International method – automated segmented flow (e.g. Skalar) International method – manual	≤60 >60	RMean	5.5 Robust SD	DU, dry basis	0
	Dumas	All	RMean	0.028	%, dry basis	2
Total Nitrogen	Kjeldahl	All	RMean	0.028	%, dry basis	2
	NIR	All	RMean (Dumas)	0.04	%, dry basis	2
Friability	Friability meter	All	Expert	1.7	%	1
Homogeneity	Friability meter	95 to 100	RMean	0.5	%	1
Glassy (Whole) Corns	Friability meter	All	RMean	0.4	%	1
Partly Unmodified Grains	Friability meter	All	RMean	0.52	%	1
Sieving Test <2.20mm	Steinecker/Glasblaserei Sortimat Other	All	RMean (Steinecker)	0.3	%	1
Sieving Test 2.20 to 2.50mm	Steinecker/Glasblaserei (EBC Fraction III) Sortimat Other	All	RMean (Steinecker)	0.5	%	1
Sieving Test 2.50 to 2.80mm	Steinecker/Glasblaserei (EBC Fraction II) Sortimat Other	All	RMean (Steinecker)	1.5	%	1

Analyte	Method	Range	AV	SDPA	Units	DP
Sieving Test >2.80mm	Steinecker/Glasblaserei (EBC Fraction I) Sortimat Other	All	RMean (Steinecker)	1.8	%	1
EBC Reject Fraction (EBC Fraction IV plus foreign matter)	Analytica EBC	All	RMean	0.5	%	1
EBC Fraction IV (<2.2mm + damaged corns from all other sieves)	Analytica EBC	All	RMean	0.5	%	1
Malt Modification	Image analyser Visual	All	RMean (all methods)	1.6	%	1
Malt Mod Homogeneity	Image analyser Visual	All	RMean (all methods)	8	%	1
DMSP (Malt) (Dimethyl Sulfide Precursor)	GC-FPD GC-FID	All	RMean	20% of AV (min 0.6)	mg/kg, as is basis	1
Free DMS (Malt) (Dimethyl Sulfide)	GC-FPD GC-FID	≤8 >8	RMean	1.0 Robust SD	mg/kg, as is basis	1
Total DMS (Malt) (Dimethyl Sulfide)	GC-FPD GC-FID	≤10 >10	RMean	1.2 Robust SD	mg/kg, as is basis	1
DMSP EBC Wort (Dimethyl Sulphide Precursor)	GC-FPD GC-FID	≤10 >10	RMean	0.5 Robust SD	mg/kg, as is basis	1
Residual Sulfur Dioxide	GC, Monier Williams Para Rosaniline, DTNB, Ripper, Enzymatic	≤10 >10	RMean	1 2	mg/kg, as is basis	1
Cold Water Extract	Cold water extract	All	RMean	0.5	% (m/m), as is basis	1
Hartong VZ45	VZ45	All	RMean	1.5	%, dry basis	1
NDMA (Nitrosodimethylamine)	GC-TEA GC-MS	≤ 2.5 > 2.5	RMean	0.5 1.0	μg/kg, as is basis	1
Glycosidic Nitrile	Distillation	All	RMean	0.11	g CN/tonne, as is basis	1
Total Phenols	Chromatographic methods Colorimetric	All	RMean (all methods)	Robust SD	mg/kg, as is basis	1

## 2) Malt sample - EBC wort dependent tests

Analyte	Method	Range	AV	SDPA	Units	DP
Extract: 1.0mm	Paddle stirrer /Density Meter, Magnetic stirrer/Density Meter	All	RMean	0.4	%, dry basis	1
Extract: 0.2mm	Paddle stirrer /Density Meter, Magnetic stirrer/Density Meter	All	RMean	0.4	%, dry basis	1
EBC Extract Difference (0.2-1.0)	Calculation	All	RMean	0.3	%, dry basis	1
pH (EBC Wort)	pH meter	All	RMean	0.05	Units	2
Colour (EBC Wort)	Visual	≤6 >6	RMean	0.3 Robust SD	deg EBC	1
Colour (EBC Wort)	Spectrophotometric	≤6 >6	RMean	0.4 Robust SD	deg EBC	1
Boiled Wort Colour	Visual – heating plate Visual – oil bath	≤9 >9	RMean	0.8 Robust SD	deg EBC	1
(EBC Wort)	Spectrophotometer – heating plate Spectrophotometer – oil bath	≤9 >9	RMean	0.8 Robust SD	deg EBC	1
	Dumas	All	RMean	0.025	% (m/m), dry basis	2
TSN (EBC Wort)	Kjeldahl	All	RMean	0.025	% (m/m), dry basis	2
(Total Soluble Nitrogen)	Spectrophotometric	All	RMean (Kjeldahl)	0.025	% (m/m), dry basis	2
	Total UV digestion	All	RMean	0.025	% (m/m), dry basis	2
Kolbach Index	Calculation	All	RMean	1.4	%	1
FAN (Free Alpha Amino Nitrogen)	Manual Ninhydrin method Automated Ninhydrin method (e.g. Skalar) Automated other (e.g. Beermaster)	All	RMean	9	mg/L	0
Viscosity (EBC Wort)	Falling Ball Viscometer, Rotary Viscometer, Glass Capillary Viscometer	All	RMean	0.02	mPas	2
EBC Fermentability (Boiled)	Dried yeast/Density meter Fresh (brewers/distillers) Yeast/ Density meter	All	RMean (Fresh brewers/ distillers yeast)	1.0	%	1

Analyte	Method	Range	AV	SDPA	Units	DP
	Fresh (bakers) Yeast/ Density					
	meter					
Beta Glucan (EBC Wort)	Spectrophotometric – automated analyser (e.g. beermaster) Spectrophotometric – manual test kit	All	RMean	20	mg/L	0
	Fluorimetric – automated analyser (e.g. Skalar) Fluorimetric – manual injection	All	RMean	20	mg/L	0
	Enzymatic	All	RMean	20	mg/L	0
Saccharification time*	EBC 4.5.1 paragraph 8.1.5 and 8.1.6	All	Consensus	-	Minutes	0

<sup>\*</sup>Not currently included in LGC's UKAS Scope of Accreditation

## 3) Malt sample - IoB wort dependent tests

Analyte	Method	Range	AV	SDPA	Units	DP
IoB Extract 0.7mm	Paddle stirrer /Density Meter, Magnetic stirrer/Density Meter Manual stir/Density Meter	All	RMean	1.5	L deg/Kg, dry basis	1
IoB Extract 0.2mm	Paddle stirrer /Density Meter, Magnetic stirrer/Density Meter Manual stir/Density Meter	All	RMean	1.5	L deg/Kg, dry basis	1
loB Extract Difference (0.2-0.7)	Calculation	All	RMean	1	L deg/Kg, dry basis	1
IoB Soluble Extract 1.0mm	Calculation	All	RMean	0.3	% sol ext, as is	1
IoB Soluble Extract 0.7mm	Calculation	All	RMean	0.3	% sol ext, as is	1
IoB Soluble Extract 0.2mm	Calculation	All	RMean	0.3	% sol ext, as is	1
loB Sol Extract Difference (0.2-0.7)	Calculation	All	RMean	0.3	% sol ext, as is	1
loB Sol Extract Difference (0.2-1.0)	Calculation	All	RMean	0.35	% sol ext, as is	1
pH (IoB Wort)	pH meter	All	RMean	0.05	pH units	2
Colour (IoP Mort)	Visual	≤6 >6	RMean	0.3 Robust SD	deg EBC	1
Colour (IoB Wort)	Spectrophotometric	≤6 >6	RMean	0.4 Robust SD	deg EBC	1

Analyte	Method	Range	AV	SDPA	Units	DP
	Dumas	All	RMean	0.025	% (m/m), dry basis	2
TCNL (IoD Mort)	Kjeldahl	All	RMean	0.025	% (m/m), dry basis	2
TSN (IoB Wort) (Total Soluble Nitrogen)	Spectrophotometric	All	RMean (Kjeldahl)	0.025	% (m/m), dry basis	2
	Total UV digestion	All	RMean	0.025	% (m/m), dry basis	2
SNR	Dumas	All	RMean	1.5	%	1
(Soluble Nitrogen Ratio)	Kjeldahl	All	RMean	1.5	%	1
FAN (IoB Wort) (Free Alpha Amino Nitrogen)	Manual Ninhydrin method Automated Ninhydrin method (e.g. Skalar) Automated other (e.g. Beermaster)	All	RMean	8	mg/L	0
Viscosity (IoB Wort)	Falling Ball Viscometer, Rotary Viscometer, Glass Capillary Viscometer	All	RMean	0.03	mPas	2
IoB Fermentability (Unboiled 0.7mm)	Dried yeast/Density meter Fresh (brewers/distillers) Yeast/ Density meter Fresh (bakers) Yeast/ Density meter	All	RMean	0.5	%	1
Predicted Spirit Yield (as is)	NIR Calculation from fermentability	All	RMean	3	L alcohol/tonne, as is	0
	Spectrophotometric – automated analyser (e.g. beermaster) Spectrophotometric – manual test kit	All	RMean	10	mg/L	0
Beta Glucan (IoB Wort)	Fluorimetric – automated analyser (e.g. Skalar) Fluorimetric – manual injection	All	RMean	10	mg/L	0
	Enzymatic	All	RMean	10	mg/L	0

## Sample PT-MP-02: Barley Barley for chemical analysis

Participants will receive 1 x 1000ml pot of barley

Analyte	Method	Range	AV	SDPA	Units	DP
•	Oven (ISO Method)			0.2		
Moisture	Moisture meter	4 to 20	RMean (ISO oven)	0.3	% w/w	1
	NIR (Near Infrared)		,	0.2		
	Dumas		RMean	0.03	%, dry basis	2
Total Nitrogen	Kjeldahl	All	RMean	0.03	%, dry basis	2
•	NIR (Near Infrared)		RMean (Dumas)	0.03	%, dry basis	2
	Steinecker/Glasblaserei					
Sieving Test <2.20mm	Sortimat	All	RMean (Steinecker)	0.3	%	1
J	Other		,			
	Steinecker/Glasblaseri					
Sieving Test <2.25mm	Sortimat	All	RMean (Steinecker)	0.3	%	1
•	Other	¬ ' '				
Sieving Test 2.20 to	Steinecker/Glasblaserei					
	(EBC Fraction III)	Δ11	DMoon (Staingalean)	4	0/	4
2.50mm	Sortimat	All	RMean (Steinecker)	1	%	1
	Other					
Signing Toot 2.25 to	Steinecker/Glasblaserei					
Sieving Test 2.25 to 2.50mm	Sortimat	All	RMean (Steinecker)	1	%	1
2.5011111	Other					
	Steinecker/Glasblaserei	All			%	
Sieving Test >2.50mm	Sortimat		RMean (Steinecker)	1.5		1
	Other					
	Steinecker/Glasblaserei					
Sieving Test 2.50 to	(EBC Fraction II)	All	RMean (Steinecker)	3	%	1
2.80mm	Sortimat	^"	(Stelliecker)	3	/0	'
	Other					
	Steinecker/Glasblaserei					
Sieving Test >2.80mm	(EBC Fraction I)	All	PMoan (Stoinocker)	3.5	%	1
Sleving rest >2.00mm	Sortimat	All	RMean (Steinecker)	3.5	70	· •
	Other					
EBC Reject Fraction						
(EBC Fraction IV plus	Analytica EBC	All	RMean	0.5	%	1
Foreign Matter)						

MAPS Scheme Description

Analyte	Method	Range	AV	SDPA	Units	DP
EBC Fraction IV (<2.2mm + damaged corns from all other sieves)	Analytica EBC	All	RMean	0.5	%	1
Thousand Corn Weight	EBC 3.4 (Automatic corn counter) EBC 3.4 (Manual corn counter)	All	RMean	0.6	g, dry basis	1
Germinative capacity	Rapid ( $TET_M$ ) staining	90 to 100	RMean	1.5	%	0
Germinative capacity	Peroxide	90 to 100	RMean	1.5	%	0
	Aubry 5 day test	90 to 100	RMean	1	%	0
Germinative energy	Schonfeld 5 day test	80 to 100	RMean	1	%	0
	BRF 4ml Test (72hrs)	85 to 100	RMean	1.5	%	0
BRF 8ml Test	72hrs	≤90 >90	RMean	10 5	%	0
Water sensitivity	ASBC MEBAK	All	RMean	Robust SD	%	0
Hectolitre Weight	Chrondrometer NIR	All	RMean	0.7	Kg/hectolitre	1

Sample PT-MP-03: Mycotoxin Malt flour for mycotoxin analysis

Participants will receive 2 x 100g pots of malt flour (sample and blank\*)

Analyte	Method	Range	AV	SDPA	Units	DP
Ochratoxin A	ELISA Lateral flow (e.g. Charm, Neogen) HPLC	All	RMean	25% of AV	μg/kg	1
Deoxynivalenol (DON)	ELISA Lateral flow (e.g. Charm, Neogen) HPLC	All	RMean	25% of AV	μg/kg	0

<sup>\*</sup>Blank material supplied as a quality control material only (results are not required to be returned for this material).

Participants will receive 1 x 300ml pot

Analyte	Method	Range	AV	SDPA	Units	DP
Moisture	Oven Moisture meter	All	RMean	0.15	% w/w	2
Colour	EBC 5.6, Vis Col, 450g mixed mash EBC 5.10 / IoB 3.6, Vis Col, 450g boiled mash IoB 3.2.4, Vis Col, 515ml boiled mash EBC 5.6, 430nm Spec Col, 450g mixed mash EBC 5.10 / IoB 3.6, 430nm Spec Col, 450g boiled mash IoB 3.2.4, 430nm Spec Col, 515ml boiled mash	All	RMean	10% of AV	deg EBC	1

Participants will receive 1 x 300ml pot

Analyte	Method	Range	AV	SDPA	Units	DP
Moisture	Oven Moisture meter	All	RMean	0.15	% w/w	2
Colour	515ml Spec Wort Colour, (IoB 3.3, 2 hour extraction @ 65°C, 50/50 mixed mash) 515ml Visual Wort Colour (IoB 3.3, 2 hour extraction @ 65°C, 50/50 mixed mash) 450g Spec Wort Colour (IoB 3.3 / EBC 5.7, 2 hour extraction @ 65°C, 50/50 mixed mash) 450g Visual Wort Colour (IoB 3.3 / EBC 5.7, 2 hour extraction @ 65°C, 50/50 mixed mash) 450g Spec Wort Colour (EBC 5.5, 45 – 70°C extraction, 50/50 mixed mash) 450g Visual Wort Colour (EBC 5.5, 45 – 70°C extraction, 50/50 mixed mash)	All	RMean	10% of AV	deg EBC	1
Degrees of crystallisation	Visual	All	RMean	Robust SD	%	1

Sample PT-MP-05: NDMA Malt flour for NDMA analysis

Participants will receive 1 x 200g flour based matrix

Analyte	Method	Range	AV	SDPA	Units	DP
NDMA	GC-TEA	≤ 2.5	RMean	0.5	ua/ka as is basis	1
(Nitrosodimethylamine)	GC-MS	> 2.5	Riviean	1.0	μg/kg, as is basis	ı

Text written in italics is for reference purposes only and will not feature in the published report.

Sample PT-MP-06: High diastatic power malt for chemical analysis

Participants will receive 1 x 1000ml pot of high DP malt

### 1) Malt sample - High diastatic power malt dependent tests

Analyte	Method	Range	AV	SDPA	Units	DP
Moisture	Oven Method	All	RMean	0.15	% w/w	1
DPWK (Diastatic power)	Iodometric (reference method) Segmented flow (e.g. Skalar) Calculated value (from IOB)	<pre>≤350 350-500 &gt;500</pre>	RMean (lodometric)	20 25 30	WK units, dry basis	0
Diastatic Power (DP IoB)	Fehling's Ferricyanide Segmented flow (e.g. Skalar) Calculated value (from DPWK)	- ≤91 - 91-128 - >128	Calculated value	9.1 10.4 11.6	IoB, as is basis	0
Alpha Amylase	International method – automated segmented flow (e.g. Skalar) International method – manual	≤60 >60	RMean	5.5 Robust SD	DU, dry basis	0
IoB Soluble Extract 0.7mm	Calculation	All	RMean	0.3	% sol ext, as is	1
TSN (IoB Wort) (Total Soluble Nitrogen)	Dumas	All	RMean	0.025		
	Kjeldahl Spectrophotometric				% (m/m), dry basis	2
	Total UV digestion					

Analyte	Method	Range	AV	SDPA	Units	DP
FAN (IoB Wort) (Free Alpha Amino Nitrogen)	Manual Ninhydrin method Automated Ninhydrin method (e.g. Skalar) Automated other (e.g. Beermaster)	All	RMean	8	mg/L	0
Glycosidic Nitrile	Distillation	All	RMean	0.11	g CN/tonne, as is basis	1

Sample PT-MP-7: Malted wheat for chemical analysis\*

Participants will receive 1 x 1000ml pot of malted wheat

## 1) Malt sample - Malted wheat dependent tests

Analyte	Method	Range	AV	SDPA	Units	DP
Moisture	Oven Method	All	RMean	0.15	% w/w	1
DPWK (Diastatic power)	Iodometric (reference method) Segmented flow (e.g. Skalar) Calculated value (from IOB)	- ≤350 - 350-500 >500	RMean (lodometric)	20 25 30	WK units, dry basis	0
Diastatic Power (DP IoB)	Fehling's Ferricyanide Segmented flow (e.g. Skalar) Calculated value (from DPWK)	- ≤91 - 91-128 - >128	RMean	9.1 10.4 11.6	IoB, as is basis	0
Alpha Amylase	International method – automated segmented flow (e.g. Skalar) International method – manual	≤60 >60	RMean	5.5 Robust SD	DU, dry basis	0
	Dumas	All	RMean	0.16	%, dry basis**	2
Drotoin	Kjeldahl	All	RMean	0.16	%, dry basis**	2
Protein	NIR	All	RMean (Dumas)	0.23	%, dry basis**	2
Extract: 0.2mm (EBC)	Paddle stirrer /Density Meter, Magnetic stirrer/Density Meter	All	RMean	0.4	%, dry basis	1
pH (EBC Wort)	pH meter	All	RMean	0.05	pH units	2
Colour (EBC Wort)	Visual	≤6 >6	RMean	0.3 Robust SD	deg EBC	1
	Spectrophotometric	≤6 >6	RMean	0.4 Robust SD	deg EBC	1

Analyte	Method	Range	AV	SDPA	Units	DP
Boiled Wort Colour (EBC Wort)	Visual – heating plate Visual – oil bath	≤9 >9	RMean	0.8 Robust SD	deg EBC	1
	Spectrophotometer – heating plate Spectrophotometer – oil bath	≤9 >9	RMean	0.8 Robust SD	deg EBC	1
	Dumas	All	RMean	0.025	% (m/m), dry basis	2
TSN (EBC Wort) (Total Soluble Nitrogen)	Kjeldahl	All	RMean	0.025	% (m/m), dry basis	2
,	Spectrophotometric	All	RMean (Kjeldahl)	0.025	% (m/m), dry basis	2
Kolbach Index	Dumas	All	RMean	1.4	%	1
	Kjeldahl	All	RMean	1.4	%	1
FAN (EBC wort) (Free Alpha Amino Nitrogen)	Manual Ninhydrin method Automated Ninhydrin method (e.g. Skalar) Automated other (e.g. Beermaster)	All	RMean	9	mg/L	0
Viscosity (EBC Wort)	Falling Ball Viscometer, Rotary Viscometer, Glass Capillary Viscometer	All	RMean	0.02	mPas	2
EBC Fermentability (Boiled)	Dried yeast/Density meter Fresh (brewers/distillers) Yeast/ Density meter Fresh (bakers) Yeast/ Density meter	All	RMean (Fresh brewers/distille rs yeast)	1.4	%	1
Beta Glucan (EBC Wort)	Spectrophotometric – automated analyser (e.g. beermaster) Spectrophotometric – manual test kit	All	RMean	20	mg/L	0
	Fluorimetric – automated analyser (e.g. Skalar) Fluorimetric – manual injection	All	RMean	20	mg/L	0
	Enzymatic	All	RMean	20	mg/L	0
IoB Extract 0.7mm	Paddle stirrer /Density Meter, Magnetic stirrer/Density Meter Manual stir/Density Meter	All	RMean	1.5	L deg/Kg, dry basis	1
pH (IoB Wort)	pH meter	All	RMean	0.05	pH units	2
Colour (IoB Wort)	Visual	≤6	RMean	0.3	deg EBC	1

Analyte	Method	Range	AV	SDPA	Units	DP
		>6		Robust SD		
	Spectrophotometric	≤6 >6	RMean	0.4 Robust SD	deg EBC	1
	Dumas	All	RMean	0.025	% (m/m), dry basis	2
TSN (IoB Wort)	Kjeldahl	All	RMean	0.025	% (m/m), dry basis	2
(Total Soluble Nitrogen)	Spectrophotometric	All	RMean (Kjeldahl)	0.025	% (m/m), dry basis	2
	Total UV digestion	All	RMean	0.025	% (m/m), dry basis	2
SNR	Dumas	All	RMean	1.5	%	1
(Soluble Nitrogen Ratio)	Kjeldahl	All	RMean	1.5	%	1
FAN (IoB Wort) (Free Alpha Amino Nitrogen)	Manual Ninhydrin method Automated Ninhydrin method (e.g. Skalar) Automated other (e.g. Beermaster)	All	RMean	8	mg/L	0
Viscosity (IoB Wort)	Falling Ball Viscometer, Rotary Viscometer, Glass Capillary Viscometer	All	RMean	0.03	mPas	2
Beta Glucan (IoB Wort)	Spectrophotometric – automated analyser (e.g. beermaster) Spectrophotometric – manual test kit	All	RMean	10	mg/L	0
	Fluorimetric – automated analyser (e.g. Skalar) Fluorimetric – manual injection	All	RMean	10	mg/L	0
	Enzymatic	All	RMean	10	mg/L	0

<sup>\*</sup>Not currently included in LGC's UKAS Scope of Accreditation
\*\*Results may also be reported expressed as Nitrogen. These values will be converted to protein using the factor 5.7